

FIT TEST REPORT

5/8/2020

ID NUMBER	13		
LAST NAME	PERSON 02	CUSTOM1	
FIRST NAME	TEST	CUSTOM2	
COMPANY	TU DELFT	CUSTOM3	
LOCATION		CUSTOM4	
TEST DATE	5/8/2020 14:45	PORTACOUNT S/N	8038200910
DUE DATE	5/8/2021	N95 COMPANION	N
RESPIRATOR	AIR-WAVE.ORG S/M NIEUW [100]	PROTOCOL	OSHA 29CFR1910.134
MANUFACTURER	DECATHLON SUBEA	PASS LEVEL	100
MODEL	EASYBREATH 500		
MASK STYLE		APPROVAL	
MASK SIZE	S/M	EFFICIENCY<99%	False

<u>EXERCISE</u>	<u>DURATION (sec.)</u>	<u>FIT FACTOR</u>	<u>PASS</u>
NORMAL BREATHING	64	550	Y
DEEP BREATHING	64	575	Y
HEAD SIDE TO SIDE	64	640	Y
HEAD UP AND DOWN	64	689	Y
TALKING	64	640	Y
GRIMACE	15	Excl.	
BENDING OVER	64	670	Y
NORMAL BREATHING	64	614	Y
OVERALL FF		622	Y

FIT TEST OPERATOR _____ **DATE** _____
WEINZIERL

NAME _____ **DATE** _____
TEST PERSON 02

Note:

Respirator Fit Test Card

Name: TEST PERSON 02 Test Date: 5/8/2020

ID: 13 Next Test Date: 5/8/2021

Respirator

Mfg: DECATHLON SUBEA Overall FF: 622
Model: EASYBREATH 500 FF Pass Level: 100
Style: NIEUW Pass: Y
Size: S/M Operator: WEINZIERL

Results

Protocol: OSHA 29CFR1910.134

Fit Test Method: QNFT using TSI PortaCount

*** www.air-wave.org ***

***Note Stichting Air Wave.org:**

Due to GDPR regulations Stichting Air-Wave.org removed the name of the test person. This name is available on request by sending an e-mail to communicatie@air-wave.org

This document is the property of Stichting Air-Wave.org and shall not be disclosed or forwarded without prior written consent. © All copyrights reserved - 2020.